

Quality certificate
Technical data for Validation
of LCK114 (150-1000 mg/l
COD)

Quality certificate

Technical data for cuvette test LCK114

Sensitivity	0.0005 Abs./(mg/l)
Ordinate intersect	0.048 Abs.
Residual standard deviation	0.0037 Abs.
Method variation coefficient	1.33 %
Method standard deviation	7.3 mg/l
Confidence intervall (95%)	± 17.7 mg/l

The technical data for cuvette test LCK114 were determined in conformity with ISO 8466-1 and DIN 38402 A51 „Calibration of analysis methods“.


The series of the smallest and largest calibration standards exhibit normal distribution and are outlier- and trend-free.

The calibration gives a linear function.

Technical data in conformity with DIN 32645

Detection limit	4.4 mg/l
Quantitation limit	13.2 mg/l

The detection and the quantitation limits were determined in conformity with DIN 32645.



Result	Confidence intervall (95%)
200 mg/l	± 18.8 mg/l
400 mg/l	± 17.8 mg/l
600 mg/l	± 17.7 mg/l
800 mg/l	± 18.3 mg/l
1000 mg/l	± 19.5 mg/l

HACH LANGE GmbH
Quality Management

Dr. Ralf Kloos